Statement of Memory Volatility

Product and Manufacturer			
Models: 393/393 FC CAT III 1500V TRMS Clamp Meter	**Manufacturer:** Fluke Corporation		
	Address: 6920 Seaway Blvd.		
	City: Everett	**State:** WA	**Zip:** 98203	**Phone:** 800-283-5853

Volatile Memory

Does the item contain volatile memory (i.e., memory whose contents are lost when power is removed)?

☒ Yes ☐ No (If the answer is 'Yes', provide the following information for each type of memory.)

Memory Type (SRAM, DRAM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
MSP430F47187 Internal SRAM	8 KB	☐ Yes ☒ No	Measurement and control	Power off
CC2541F256 Internal SRAM	8 KB	☐ Yes ☒ No	Communication and control	Power off

Non-Volatile Memory

Does the item contain non-volatile memory (i.e., memory whose contents are retained when power is removed)?

☒ Yes ☐ No (If the answer is 'Yes', provide the following information for each type of memory.)

Memory Type (BBRAM, Flash, EEPROM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
MSP430F47187 Internal Flash	112 kB	☐ Yes ☒ No	Measurement and control	NA
CC2541F256 Internal Flash	256 kB	☐ Yes ☒ No	Communication and control	NA
AT25DF081A Internal Flash	1 MB	☒ Yes ☐ No	Test result storage	Clear with keypad or APP

Media

Does the item contain media storage capability (i.e., removable or non-removable disk drives, memory cards, etc.)?

☐ Yes ☒ No (If the answer is 'Yes', provide the following information for each type of memory.)

	Storage Type (Disk, Tape, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
			☐ Yes ☐ No		

Additional Information

NA

Fluke Representative Information

Name:	Title:	Office Phone:	Fax/Email:
Susan Isaac	Product Manager	1-425-347-6100	susan.isaac@fluke.com